


Quality Control Certificate



dsDNA-Ab ELISA

REF RE75201

LOT 25070141

 2026-10-31
YYYY-MM-DD

Final Release Results:

	Conc. IU/mL	Measured value OD	Acceptable range		
			lower limit OD	target value OD	upper limit OD
Calibrator A	0	0.055			< 0.100
Calibrator F	250	2.250	> 1.000		
CONTROL -	-	0.070			≤ 0.300
CONTROL +	-	0.882	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value IU/mL	Acceptable range		
		lower limit IU/mL	target value IU/mL	upper limit IU/mL
CONTROL -	1.4			≤ 10.0
CONTROL +	82	51	80	109

Factor for qualitative assay evaluation	0.63
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2025-08-01

This document has been produced electronically and is valid without a signature.