


# Quality Control Certificate



## dsDNA-Ab ELISA

REF RE75201

LOT 26040061

 2027-06-30  
YYYY-MM-DD

### Final Release Results:

	Conc. IU/mL	Measured value OD	Acceptable range		
			lower limit OD	target value OD	upper limit OD
Calibrator A	0	0.053			< 0.100
Calibrator F	250	2.052	> 1.000		
CONTROL -	-	0.080			≤ 0.300
CONTROL +	-	1.115	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value IU/mL	Acceptable range		
		lower limit IU/mL	target value IU/mL	upper limit IU/mL
CONTROL -	1.9			≤ 10.0
CONTROL +	85	51	80	109

Factor for qualitative assay evaluation	0.63
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

# Accepted

Validation date: 2026-04-15

This document has been produced electronically and is valid without a signature.