


# Quality Control Certificate



## U1-RNP-Ab ELISA

REF RE75211

LOT 25030075

 2026-05-31  
YYYY-MM-DD

### Final Release Results:

	Conc.	Measured value	Acceptable range		
	U/mL		lower limit	target value	upper limit
		OD	OD	OD	OD
CAL A	0	0.049			< 0.100
CAL F	60	1.945	> 1.000		
CONTROL -	-	0.090			≤ 0.300
CONTROL +	-	0.926	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value	Acceptable range		
	U/mL	lower limit	target value	upper limit
		U/mL	U/mL	U/mL
CONTROL -	0.5			≤ 2.0
CONTROL +	13	10	16	22

Factor for qualitative assay evaluation	0.37
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

# Accepted

Validation date: 2025-03-13

This document has been produced electronically and is valid without a signature.