


Quality Control Certificate



U1-RNP-Ab ELISA

REF RE75211

LOT 26050068

 2027-07-31
YYYY-MM-DD

Final Release Results:

	Conc.	Measured value	Acceptable range		
	U/mL		lower limit	target value	upper limit
		OD	OD	OD	OD
CAL A	0	0.051			< 0.100
CAL F	60	2.371	> 1.000		
CONTROL -	-	0.103			≤ 0.300
CONTROL +	-	1.204	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value	Acceptable range		
		lower limit	target value	upper limit
U/mL	U/mL	U/mL	U/mL	
CONTROL -	0.6			≤ 2.0
CONTROL +	18	10	16	22

Factor for qualitative assay evaluation	0.37
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2026-05-07

This document has been produced electronically and is valid without a signature.