

Quality Control Certificate



Sm-Ab ELISA

REF RE75221

LOT 25030076

 2026-05-31
YYYY-MM-DD

Final Release Results:

	Conc. U/mL	Measured value OD	Acceptable range		
			lower limit OD	target value OD	upper limit OD
CAL A	0	0.053			< 0.100
CAL F	40	2.249	> 1.000		
CONTROL -	-	0.119			≤ 0.300
CONTROL +	-	1.097	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value U/mL	Acceptable range		
		lower limit U/mL	target value U/mL	upper limit U/mL
CONTROL -	0.4			≤ 2.0
CONTROL +	10	6	10	14

Factor for qualitative assay evaluation	0.52
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2025-03-13

This document has been produced electronically and is valid without a signature.