


Quality Control Certificate



Sm-Ab ELISA

REF RE75221

LOT 26040062

 2027-06-30
YYYY-MM-DD

Final Release Results:

	Conc.	Measured value	Acceptable range		
	U/mL		lower limit	target value	upper limit
		OD	OD	OD	OD
CAL A	0	0.049			< 0.100
CAL F	40	2.149	> 1.000		
CONTROL -	-	0.110			≤ 0.300
CONTROL +	-	1.031	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value	Acceptable range		
	U/mL	lower limit	target value	upper limit
	U/mL	U/mL	U/mL	U/mL
CONTROL -	0.4			≤ 2.0
CONTROL +	10	6	10	14

Factor for qualitative assay evaluation	0.52
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2026-04-17

This document has been produced electronically and is valid without a signature.