


# Quality Control Certificate



## SS-A/Ro-Ab ELISA

REF RE75231

LOT 26050090

 2027-07-31  
YYYY-MM-DD

### Final Release Results:

	Conc.	Measured value	Acceptable range		
	U/mL		lower limit	target value	upper limit
		OD	OD	OD	OD
CAL A	0	0.050			< 0.100
CAL F	70	2.230	> 1.000		
CONTROL -	-	0.128			≤ 0.300
CONTROL +	-	1.328	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value	Acceptable range		
	U/mL	lower limit	target value	upper limit
	U/mL	U/mL	U/mL	U/mL
CONTROL -	0.4			≤ 2.0
CONTROL +	15	10	16	22

Factor for qualitative assay evaluation	0.34
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

# Accepted

Validation date: 2026-06-11

This document has been produced electronically and is valid without a signature.