


Quality Control Certificate



SS-B/La-Ab ELISA

REF RE75241

LOT 26040058

 2027-05-31
YYYY-MM-DD

Final Release Results:

	Conc.	Measured value	Acceptable range		
	U/mL		lower limit	target value	upper limit
		OD	OD	OD	OD
CAL A	0	0.046			< 0.100
CAL F	60	1.958	> 1.000		
CONTROL -	-	0.095			≤ 0.300
CONTROL +	-	0.961	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value	Acceptable range		
	U/mL	lower limit	target value	upper limit
		U/mL	U/mL	U/mL
CONTROL -	0.5			≤ 2.0
CONTROL +	15	10	16	22

Factor for qualitative assay evaluation	0.73
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2026-04-15

This document has been produced electronically and is valid without a signature.