

Quality Control Certificate



SS-B/La-Ab ELISA

REF RE75241

LOT 26070089

2027-10-31
YYYY-MM-DD

Final Release Results:

	Conc.	Measured value	Acceptable range		
	U/mL	OD	lower limit OD	target value OD	upper limit OD
CAL A	0	0.051	> 1.000		
CAL F	60	2.307			
CONTROL -	-	0.093	≥ 0.500		
CONTROL +	-	1.092			

	Validation criteria (manual)			
	Measured value	Acceptable range		
	U/mL	lower limit U/mL	target value U/mL	upper limit U/mL
CONTROL -	0.4	10	16	≤ 2.0
CONTROL +	17			22

Factor for qualitative assay evaluation	0.73
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-07-31

Accepted

This document has been produced electronically and is valid without a signature.