

# Quality Control Certificate



## Sci-70-Ab ELISA

REF RE75251

LOT 26050070

 2027-07-31  
YYYY-MM-DD

### Final Release Results:

|           | Conc. | Measured value | Acceptable range |              |             |
|-----------|-------|----------------|------------------|--------------|-------------|
|           | U/mL  |                | lower limit      | target value | upper limit |
|           |       | OD             | OD               | OD           | OD          |
| CAL A     | 0     | 0.052          |                  |              | < 0.100     |
| CAL F     | 60    | 2.161          | > 1.000          |              |             |
| CONTROL - | -     | 0.077          |                  |              | ≤ 0.300     |
| CONTROL + | -     | 1.051          | ≥ 0.500          |              | ≤ 1.600     |

|           | Validation criteria (manual) |                  |              |             |
|-----------|------------------------------|------------------|--------------|-------------|
|           | Measured value               | Acceptable range |              |             |
|           |                              | lower limit      | target value | upper limit |
|           | U/mL                         | U/mL             | U/mL         | U/mL        |
| CONTROL - | 0.2                          |                  |              | ≤ 2.0       |
| CONTROL + | 15                           | 10               | 15           | 20          |

|   |      |
|---|------|
| Factor for qualitative assay evaluation | 0.45 |
|---|------|

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

# Accepted

Validation date: 2026-05-07

This document has been produced electronically and is valid without a signature.