

Quality Control Certificate



Sci-70-Ab ELISA

REF RE75251

LOT 2606136

 2027-11-30
YYYY-MM-DD

Final Release Results:

	Conc.	Measured value	Acceptable range		
	U/mL	OD	lower limit OD	target value OD	upper limit OD
CAL A	0	0.043			< 0.100
CAL F	60	2.000	> 1.000		
CONTROL -	-	0.076			≤ 0.300
CONTROL +	-	1.135	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value U/mL	lower limit U/mL	target value U/mL	upper limit U/mL
CONTROL -	0.3			≤ 2.0
CONTROL +	16	10	15	20

Factor for qualitative assay evaluation	0.45
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-08-25

Accepted

This document has been produced electronically and is valid without a signature.