


# Quality Control Certificate



## Jo-1-Ab ELISA

REF RE75271

LOT 25060165

 2026-07-31  
YYYY-MM-DD

### Final Release Results:

|           | Conc.<br>U/mL | Measured value<br>OD | Acceptable range  |                    |                   |
|-----------|---------------|----------------------|-------------------|--------------------|-------------------|
|           |               |                      | lower limit<br>OD | target value<br>OD | upper limit<br>OD |
| CAL A     | 0             | 0.095                |                   |                    | < 0.100           |
| CAL F     | 100           | 2.143                | > 1.000           |                    |                   |
| CONTROL - | -             | 0.186                |                   |                    | ≤ 0.300           |
| CONTROL + | -             | 1.275                | ≥ 0.500           |                    | ≤ 1.600           |

|           | Validation criteria (manual) |                     |                      |                     |
|-----------|------------------------------|---------------------|----------------------|---------------------|
|           | Measured value<br>U/mL       | Acceptable range    |                      |                     |
|           |                              | lower limit<br>U/mL | target value<br>U/mL | upper limit<br>U/mL |
| CONTROL - | 0.9                          |                     |                      | ≤ 2.0               |
| CONTROL + | 26                           | 17                  | 27                   | 37                  |

|   |      |
|---|------|
| Factor for qualitative assay evaluation | 0.28 |
|---|------|

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

# Accepted

Validation date: 2025-08-04

This document has been produced electronically and is valid without a signature.