

# Quality Control Certificate



## ANA Profile 8 ELISA

REF	<b>RE75401</b>	LOT	<b>26050069</b>		<b>2027-01-31</b> YYYY-MM-DD
-----	----------------	-----	-----------------	--	---------------------------------

### Final Release Results:

Parameter	Factor	Positive control (Cal.) OD		Negative control OD	
		Acceptable range	Result	Acceptable range	Result
LA: dsDNA	0.47	0.500 - 2.000	0.806	0.020 - 0.180	0.050
LB: RNP	0.32	0.500 - 2.000	0.799	0.020 - 0.180	0.052
LC: Sm	0.43	0.500 - 2.000	0.945	0.020 - 0.180	0.055
LD: SS-A/Ro	0.50	0.500 - 2.000	1.301	0.020 - 0.180	0.064
LE: SS-B/La	0.62	0.500 - 2.000	0.829	0.020 - 0.180	0.051
LF: Scl-70	0.41	0.500 - 2.000	0.825	0.020 - 0.180	0.052
LG: CENP-B	0.36	0.500 - 2.000	0.912	0.020 - 0.180	0.058
LH: Jo-1	0.40	0.500 - 2.000	0.895	0.020 - 0.180	0.054

L= Line

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

**Accepted**

Validation date: 2026-05-08

This document has been produced electronically and is valid without a signature.



IBL International GmbH  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.:  
E-MAIL:  
WEB:

+ 49 (0) 40 532891 -0 Fax: -11  
[ibl@tecan.com](mailto:ibl@tecan.com)  
[www.tecan.com/ibl](http://www.tecan.com/ibl)