

Quality Control Certificate



ANA Profile 8 ELISA

REF	RE75401	LOT	26060074		2027-04-30 YYYY-MM-DD
-----	---------	-----	----------	--	--------------------------

Final Release Results:

Parameter	Factor	Positive control (Cal.) OD		Negative control OD	
		Acceptable range	Result	Acceptable range	Result
LA: dsDNA	0.47	0.500 - 2.000	0.505	0.020 - 0.180	0.053
LB: RNP	0.32	0.500 - 2.000	0.883	0.020 - 0.180	0.064
LC: Sm	0.43	0.500 - 2.000	1.024	0.020 - 0.180	0.070
LD: SS-A/Ro	0.50	0.500 - 2.000	1.670	0.020 - 0.180	0.094
LE: SS-B/La	0.62	0.500 - 2.000	0.859	0.020 - 0.180	0.062
LF: Scl-70	0.41	0.500 - 2.000	0.671	0.020 - 0.180	0.057
LG: CENP-B	0.36	0.500 - 2.000	1.214	0.020 - 0.180	0.078
LH: Jo-1	0.40	0.500 - 2.000	1.009	0.020 - 0.180	0.067

L= Line

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2026-07-16

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.:
E-MAIL:
WEB:

+ 49 (0) 40 532891 -0 Fax: -11
ibl@tecan.com
www.tecan.com/ibl