

Quality Control Certificate



ENA Screen 6 ELISA

REF RE75411	LOT 26030171	2027-08-31 YYYY-MM-DD
--------------------	---------------------	---------------------------------

Final Release Results:

Controls	Result OD	Acceptable range OD
Positive control	1.274	0.600 - 1.700
Negative control	0.080	≤ 0.180

Factor for qualitative assay evaluation	0.42
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-05-21

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.:
E-MAIL:
WEB:

+ 49 (0) 40 532891 -0 Fax: -11
ibl@tecan.com
www.tecan.com/ibl