

Quality Control Certificate



ANA Screen 8 ELISA

REF RE75421	LOT 25030078	2026-05-31 YYYY-MM-DD
--------------------	---------------------	---------------------------------

Final Release Results:

Controls	Result OD	Acceptable range OD
Positive control	0.930	0.600 - 1.700
Negative control	0.064	≤ 0.200
Factor for qualitative assay evaluation		0.30

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-03-13

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.:
E-MAIL:
WEB:

+ 49 (0) 40 532891 -0 Fax: -11
ibl@tecan.com
www.tecan.com/ibl