

Quality Control Certificate



ANA Screen 8 ELISA

REF RE75421	LOT 26010074	2027-04-30 YYYY-MM-DD
--------------------	---------------------	---------------------------------

Final Release Results:

Controls	Result OD	Acceptable range OD
Positive control	0.937	0.600 - 1.700
Negative control	0.062	≤ 0.200
Factor for qualitative assay evaluation		0.30

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-02-10

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl