

# Quality Control Certificate



## ANA Screen 8 ELISA

<b>REF</b> RE75421	<b>LOT</b> 26040086	<b>2027-06-30</b> YYYY-MM-DD
--------------------	---------------------	---------------------------------

### Final Release Results:

Controls	Result OD	Acceptable range OD
Positive control	0.963	0.600 - 1.700
Negative control	0.066	≤ 0.200
Factor for qualitative assay evaluation		0.30

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-05-07

**Accepted**

This document has been produced electronically and is valid without a signature.



**IBL International GmbH**  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)