


Quality Control Certificate



Thyroglobulin-Ab ELISA

REF RE75501

LOT 26030172

 2027-11-30
YYYY-MM-DD

Final Release Results:

	Conc. IU/mL	Measured value OD	Acceptable range		
			lower limit OD	target value OD	upper limit OD
Calibrator A	0	0.059			< 0.100
Calibrator F	3000	2.188	> 1.000		
CONTROL -	-	0.099			< 0.300
CONTROL +	-	1.141	> 0.500		< 1.600

	Validation criteria (manual)			
	Measured value IU/mL	Acceptable range		
		lower limit IU/mL	target value IU/mL	upper limit IU/mL
CONTROL -	9.5			< 25.0
CONTROL +	528	384	600	816

Factor for qualitative assay evaluation	0.29
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2026-06-11

This document has been produced electronically and is valid without a signature.