


Quality Control Certificate



Thyroid Peroxidase-Ab ELISA

REF RE75511

LOT 25090225

 2027-03-31
YYYY-MM-DD

Final Release Results:

	Conc. IU/mL	Measured value OD	Acceptable range		
			lower limit OD	target value OD	upper limit OD
Calibrator A	0	0.058			< 0.100
Calibrator F	3000	2.415	> 1.000		
CONTROL -	-	0.164			≤ 0.300
CONTROL +	-	1.683	≥ 0.800		≤ 1.900

	Validation criteria (manual)			
	Measured value IU/mL	Acceptable range		
		lower limit IU/mL	target value IU/mL	upper limit IU/mL
CONTROL -	12.7			≤ 25.0
CONTROL +	462	288	450	612

Factor for qualitative assay evaluation	0.39
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2025-10-10

This document has been produced electronically and is valid without a signature.