


Quality Control Certificate



Myeloperoxidase-Ab (P-ANCA) ELISA

REF RE75601

LOT 26010089

 2027-03-31
YYYY-MM-DD

Final Release Results:

	Conc.	Measured value	Acceptable range		
	U/mL		lower limit	target value	upper limit
		OD	OD	OD	OD
Calibrator A	0	0.051			< 0.100
Calibrator F	100	2.097	> 1.000		
CONTROL -	-	0.146			≤ 0.300
CONTROL +	-	1.090	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value	Acceptable range		
		lower limit	target value	upper limit
U/mL	U/mL	U/mL	U/mL	
CONTROL -	1.5			≤ 4.0
CONTROL +	28	19	30	41

Factor for qualitative assay evaluation	0.47
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2026-01-22

This document has been produced electronically and is valid without a signature.