

Quality Control Certificate



Myeloperoxidase-Ab (P-ANCA) ELISA

REF RE75601

LOT 26080119

 2028-02-29
YYYY-MM-DD

Final Release Results:

	Conc.	Measured value	Acceptable range		
	U/mL	OD	lower limit OD	target value OD	upper limit OD
Calibrator A	0	0.039	> 1.000		
Calibrator F	100	2.450			
CONTROL -	-	0.155	≥ 0.500		
CONTROL +	-	1.434			

	Validation criteria (manual)			
	Measured value	Acceptable range		
	U/mL	lower limit U/mL	target value U/mL	upper limit U/mL
CONTROL -	1.6	19	30	≤ 4.0
CONTROL +	33			41

Factor for qualitative assay evaluation	0.47
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-09-24

Accepted

This document has been produced electronically and is valid without a signature.