


# Quality Control Certificate



## Modified Gliadin Peptide IgA ELISA

REF RE75741

LOT 25120094

 2027-06-30  
YYYY-MM-DD

### Final Release Results:

	Conc. U/mL	Measured value OD	Acceptable range		
			lower limit OD	target value OD	upper limit OD
Calibrator A	0	0.051			< 0.100
Calibrator F	100	2.218	> 1.000		
CONTROL -	-	0.106			≤ 0.300
CONTROL +	-	1.021	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value U/mL	Acceptable range		
		lower limit U/mL	target value U/mL	upper limit U/mL
CONTROL -	2.0			≤ 5.0
CONTROL +	37	23	36	49

Factor for qualitative assay evaluation	0.28
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

# Accepted

Validation date: 2026-02-10

This document has been produced electronically and is valid without a signature.