


Quality Control Certificate



Modified Gliadin Peptide IgA ELISA

REF RE75741

LOT 26030173

 2027-10-31
YYYY-MM-DD

Final Release Results:

	Conc.	Measured value	Acceptable range		
	U/mL		lower limit OD	target value OD	upper limit OD
Calibrator A	0	0.051			< 0.100
Calibrator F	100	2.258	> 1.000		
CONTROL -	-	0.107			≤ 0.300
CONTROL +	-	0.978	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value	Acceptable range		
		lower limit	target value	upper limit
	U/mL	U/mL	U/mL	U/mL
CONTROL -	1.9			≤ 5.0
CONTROL +	37	23	36	49

Factor for qualitative assay evaluation	0.28
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2026-05-21

This document has been produced electronically and is valid without a signature.