


Quality Control Certificate



Modified Gliadin Peptide IgG ELISA

REF RE75751

LOT 25020148

 2026-07-31
YYYY-MM-DD

Final Release Results:

	Conc.	Measured value	Acceptable range		
	U/mL		lower limit	target value	upper limit
		OD	OD	OD	OD
Calibrator A	0	0.057			< 0.100
Calibrator F	100	2.036	> 1.000		
CONTROL -	-	0.126			≤ 0.300
CONTROL +	-	1.060	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value	Acceptable range		
	U/mL	lower limit	target value	upper limit
		U/mL	U/mL	U/mL
CONTROL -	1.8			≤ 5.0
CONTROL +	32	23	36	49

Factor for qualitative assay evaluation	0.31
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2025-03-13

This document has been produced electronically and is valid without a signature.