

Quality Control Certificate



Phenylalanine (PKU) neonatal Screening Assay

REF RE80015 & RE80019

LOT NP260

2027-01-31
YYYY-MM-DD

Final Release Results:

	Concentration mg/dL	Measured value OD	Range OD	OD/ODmax
CAL A	1.80	0.070	-	23%
CAL B	3.30	0.089	-	29%
CAL C	6.50	0.130	-	43%
CAL D	11.50	0.212	-	70%
CAL E	17.00	0.304	-	100%

	Concentration found mg/dL	Acceptable range		
		lower limit mg/dL	target value mg/dL	upper limit mg/dL
CONTROL 1	2.8	2.4	3.1	3.7
CONTROL 2	4.8	4.2	5.2	6.2

Kit Composition:

REF		LOT	
KENP655	CONTROL 1	309	2027-01
KENP665	CONTROL 2	309	2027-01
KENP545	LYO ENZ	301	2027-06
KENP755/9	ENZDIL	295	2027-03
KENP615	CAL A-E	308	2027-01
KENP775/9	DYE	294	2027-03
KENP765/9	SUBS	290	2027-03

Valid edition of the instructions for use: 2024-11

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2026-07-07

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl