

Quality Control Certificate



alpha-Amylase Saliva Assay

REF RE80111

LOT EAA137



2028-04-30

YYYY-MM-DD

Final Release Results:

	Concentration	Measured value		Delta OD	Acceptabel range
	U/mL	OD _{3min}	OD _{8min}		OD _{8min}
CAL A	320	0.892	2.141	1.250	< 2.900
CAL B	160	0.428	1.038	0.611	-
CAL C	80	0.239	0.562	0.324	-
CAL D	40.0	0.133	0.297	0.164	-
CAL E	0	0.028	0.028	0.000	-

	Concentration found	Acceptable range		
	U/mL	lower limit U/mL	target value U/mL	upper limit U/mL
CONTROL 1	50.3	28.3	47.1	65.9
CONTROL 2	149	84	132	179

Kit Composition:

REF		LOT	
KEAA651	LYO CONTROL 1	K813	2028-05
KEAA661	LYO CONTROL 2	K814	2028-05
KEAA511	MTP	K778	2028-05
KEAA611	LYO CAL	K812	2028-05
KEAA751	CONC DILBUF	K779	2028-05
KEAA771	SUBS	K797	2028-04

Valid edition of the instructions for use: 2014-04

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-07-20

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl